

**MASSACHUSETTS INSTITUTE OF TECHNOLOGY
HAYSTACK OBSERVATORY
WESTFORD, MASSACHUSETTS 01886**

August 26, 2008

*Telephone: 781-981-5407
Fax: 781-981-0590*

To: VSRT Group
From: Alan E.E. Rogers
Subject: System noise in Ozone spectrometer

The noise performance of the ozone spectrometer is measured by placing an absorber over the feed and measuring the increase in total power. The “Y” factor or ratio between the absorber in place and with the absorber removed is

$$(T_{LNA} + T_{amb}) / (T_{LNA} + T_{atmos} + T_{CMB} + T_{spill})$$

where T_{LNA} = LNA noise ~0.3 dB ~ 20K
 T_{amp} = air temperature ~ 300K
 T_{atmos} = temperature due to atmosphere absorption at 11.072 GHz 8° elevation
depends on weather ~ 35 K.
 T_{CMB} = Cosmic Microwave Background 3K
 T_{spill} = feed spill-over ~40 K

Using the numbers above we expect

$$Y = (20 + 300) / (20 + 35 + 3 + 40) \approx 5.1 \text{ dB}$$

In practice the Y factor of the “Ridge” system in the Moran building is 4.6 ± 0.3 dB and the “CHS” system is 4.0 ± 0.3 dB. The “Ridge” system uses a 75 cm dish (KDS75S) with F/D of 0.5 while the CHS is a 46 cm dish with F/D of 0.59. The deeper dish with smaller F/D ratio offers some advantage in a reduced spillover.